



NS8102 Series Navigation Signal Simulator

Datasheet



Saluki Technology Inc.

The document applies to the satellite navigation signal simulator of the following model:

- NS8102 Series

Standard package and accessories:

No.	Item
1	Main Machine
2	Power cable
3	User manual
4	Bundled RF cable

Options of the NS8102 series navigation signal simulator in addition to standard accessories (up to 2 signals):

Model No.	Description
NS8102 -01	BeiDou System: B1
NS8102-02	BeiDou System: B2
NS8102-03	BeiDou System: B3
NS8102-04	GPS: L1
NS8102-05	GPS: L2
NS8102-06	GPS: L5
NS8102-07	GLONASS: L1
NS8102-08	GLONASS: L2
NS8102-09	Galileo: E1
NS8102-10	Galileo: E5a
NS8102-11	Galileo: E5b
NS8102-12	RDSS: S
NS8102-13	RDSS: L

Preface

Thank you for choosing NS8102 series satellite navigation signal simulator produced by Saluki Technology Inc.

We devote ourselves to meeting your demands, providing you high-quality measuring instrument and the best after-sales service. We persist with “superior quality and considerate service”, and are committed to offering satisfactory products and service for our clients.

Document No.

NS8102-02-01

Version

Rev02 2019.03

Saluki Technology

Document Authorization

The information contained in this document is subject to change without notice. The power to interpret the contents of and terms used in this document rests with Saluki.

Saluki Tech owns the copyright of this document which should not be modified or tampered by any organization or individual, or reproduced or transmitted for the purpose of making profit without its prior permission, otherwise Saluki will reserve the right to investigate and affix legal liability of infringement.

Product Quality Assurance

The warranty period of the product is 36 months from the date of delivery. The instrument manufacturer will repair or replace damaged parts according to the actual situation within the warranty period.

Product Quality Certificate

The product meets the indicator requirements of the document at the time of delivery. Calibration and measurement are completed by the measuring organization with qualifications specified by the state, and relevant data are provided for reference.

Quality/Settings Management

Research, development, manufacturing and testing of the product comply with the requirements of the quality and environmental management system.

Contacts

Service Tel: 886. 909 602 109

Website: www.salukitec.com

Email: sales@salukitec.com

Address: No. 367 Fuxing N Road, Taipei 105, Taiwan (R.O.C.)

Content

1. Overview.....	6
1.1. Definitions.....	6
2. Specifications.....	7
2.1. Signal Simulation Performance.....	7
2.1.1. Signal Output Frequency.....	7
2.1.2. Modulation System.....	7
2.1.3. Signal Power Control.....	7
2.1.4. Signal Scale.....	7
2.1.5. Phase Noise.....	7
2.1.6. Spurious Level.....	7
2.1.7. Harmonics Level.....	7
2.1.8. External Interface.....	8
2.2. General Information.....	8
2.3. Compliant.....	8
2.3.1. CE.....	8
2.3.2. ISO.....	8

1. Overview

Saluki NS8102 series navigation signal simulator is a low-cost test solution for satellite navigation terminal manufacturers. It is configured with 8 signal channels. NS8102 series has the satellite navigation signal analog output capability of BDS constellation (36 RNSS), GPS constellation (32), GLONASS constellation (24), Galileo constellation (24) and its mixed constellation. Optional standard Ethernet interface for easy integration into production line ATE systems and for PCBA testing of semi-finished and finished products.

NS8102 series navigation signal simulator has many features like small size, light weight, easy to carry, abundant functions etc.. It supports both wired and wireless test methods, which can be used for user machine sensitivity, signal-to-noise ratio, first positioning time, positioning accuracy, speed measurement accuracy and other indicators test. It is suitable for car navigation, navigation mobile phone, digital products (with navigation) manufacturers to test and use.

1. 1. Definitions

Specification (Spec.)

Specifications describe the performance of parameters within the warranty of the instrument. Product specifications applies under the following conditions:

- 1) Two hours storage at ambient temperature(0-40℃) followed by 30 minutes warm-up operation
- 2) Specified environmental conditions met
- 3) Instrument is within its calibration cycle.
- 4) The specification listed in the datasheet includes measurement uncertainties.

Data in this document are Spec. unless otherwise noted.

Typical (typ.)

Typical data is not guaranteed by instrument warranty. It describes additional product performance information that 80 percent of the units exhibit. Typical data only valid at 25℃. Typical performance does not include measurement uncertainty.

Nominal (nom.)

Nominal values indicate expected performance, or describe product performance that is useful in the application of the product, but are not covered by the product warranty.

2. Specifications

2. 1. Signal Simulation Performance

2. 1. 1. Signal Output Frequency

No.	Signal Type	Frequency
1	Beidou B1	1561.098MHz
2	GPS L1	1575.42MHz
3	GLONASS L1	1602.000MHz

2. 1. 2. Modulation System

Modulation Mode	BPSK
-----------------	------

2. 1. 3. Signal Power Control

No.	Parameters	Range
1	Power range	-150dBm to -60dBm
2	Resolution	Superior to 1dB
3	Absolute accuracy	Superior to 1dB

2. 1. 4. Signal Scale

Channel Number	1 or 8 channels per frequency
----------------	-------------------------------

2. 1. 5. Phase Noise

No.	Frequency Level	Noise Value
1	100Hz	-75dBc/Hz
2	1kHz	-80dBc/Hz
3	10kHz	-85dBc/Hz
4	100kHz	-90dBc/Hz

2. 1. 6. Spurious Level

Spurious Level	<-40dBc
----------------	---------

2. 1. 7. Harmonics Level

Harmonics Level	<-35dBc
-----------------	---------

2. 1. 8. External Interface

RF output port	1 N-type
----------------	----------

2. 2. General Information

Power Supply	AC power, 220V±20V
Operation Temperature	-20℃ - +50℃
Storage Temperature	-45℃ - +75℃
Size	518.4 * 482.6 * 44mm

2. 3. Compliant

2. 3. 1. CE



- EMC

Complies with the requirements of the **EC EMC** directives.

Test Standards:EN 61326

- Safety

Complies with **EC LVD** Directive.

Test Standard:EN61010-1

2. 3. 2. ISO



- Manufacturing

This instrument is manufactured in an ISO-9001 registered facility

- End of Document -